


Operations Manual

INCUB-SHAKER™ IOL

H2010 and H2012 (-E)




Benchmark 

Table of Contents	Page
1. General information and Precautions	3
1.1 Precaution related to the power cable	3
1.2 Precautions for use	3
2. Features and Specifications	4
2.1 Features	4
2.2 Specifications	4
3. Operation	5
▶ Booting	5
▶ Basic view screen	5
▷ Temperature, RPM setting change	6
▷ Time setting change	6
▷ The end of setting time	7
▶ Advanced mode	7
▶ Temperature calibration	8
▶ Temperature setting (Advanced screen)	8
▷ Temperature setting change	8
▷ Adjusting the time of the temperature	9
▷ Adjusting the alarm setting	9
▷ Alarm off	10
▶ Shaker setting (Advanced screen)	10
▷ RPM setting change	10
▷ Adjusting the shaking time	10
▷ Alarm setting change	11
▷ Alarm off	11
▶ Temperature step setting	12
▷ Temperature time setting	12
▶ Shaker step setting	13
▷ RPM time setting	13
▶ Step cycle ending screen	14
4. Additional Operating Instructions	14
▶ H2012 (Cooling) model only	14
▶ Over temperature protection	14
▶ Platform installation	15
5. Warning Messages & Fuse Replacement	15
▶ Fuse replacement	15
▶ Event view	16

1. General Information and Precautions

Safety Symbols :



1.1. Precautions related to the power cable



Always allow at least 3cm between the power cable and the back wall to prevent the back wall from causing pressure on the power cable.

Always use the electrical plug that was supplied with this product.



Never make contact with the power cord with wet hands. (This can result in electric shock.)

Never use a damaged power cord or outlet.



In the event of smoke or a burning smell, immediately remove the power cord from the outlet.

1.2. Precautions for use



Do not attempt to disassemble this product. If service is required, please contact your local representative.



Never operate a flammable spray near the product. (This can result in a fire hazard.)

Always use caution when using flammable substances such as benzene, alcohol and LP gas. (Failure to do so can result in a fire hazard.)



Prevent foreign substances from contacting the door seal. (The inflow of outside air can negatively impact the temperature in chamber.)

Permissible ambient temperature range for transport: -10°C to 60°C.

ALWAYS ensure that the instrument is connected to an outlet with specifications that match those of the serial number label.

(Over-voltage, under-voltage can damage the product and result in poor performance.)

When installing the product, always allow a distance of at least 30cm from the wall.

Install the unit in the way that the power plug is easily accessible and can be easily pulled in case of danger.

Install the unit at a flat surface, free from vibration and in a well-ventilated location. (If the ground is not flat, it can cause an excessive vibration of the product.)

2. Features and Specifications

2.1. Features

1. Magnetic induction drive & Brush less DC motor provides less vibration and maintenance.
2. No interference with uniformity for cell culture by BLDC motor.
3. User friendly, intuitive LCD touch panel.
4. Moisture resistant structure by Epoxy-coated.
5. Automatic stop function Once LID opens, shaker automatically stops for safety and convenience.
6. One-piece construction with acrylic cover.
7. Transparent acrylic cover allows internal viewing without disturbance.
8. Stable and low noise shaking.
 - ◇ Brushless DC motor and beltless drive system for shaker minimizes noise and vibration.
9. Non-volatile memory for operation.
10. Easy-Platform Installation.
 - ◇ Very simple way to place holder or spring rack platform on the shaker.
 - ◇ Optional accessories available for most common laboratory vessels.

2.2. Specifications


ITEM	H2010	H2012
Temp. Range	Ambient +5°C to 60°C	Ambient -10°C to 60°C
Temp. Accuracy	±0.1°C at 37°C	
Temp. Uniformity	±1.0°C at 37°C	
Speed Range	30 ~ 300rpm	
Shaking Orbit	19mm (3/4")	
Circulation Fan	Convection Fan	
Timer	Digital / 1 min. to 99hr 59min. (1 min. increments)	
Motor	Beltless DC Magnetic Induction Drive	
Weight	130lbs / 50kg	185lbs / 84kg
Electrical	120V or 230V, 50~60Hz / 480W	120V or 230V, 50~60Hz / 1.2kW
Dimension		
Exterior	51(W) x 67(D) x 47(H) cm / 20(W) x 27(D) x 19(H) in.	51(W) x 81(D) x 56(H) cm / 20(W) x 32(D) x 22(H) in.
Internal	49.5(W) x 49.5(D) x 28(H) cm / 19.5(W) x 19.5(D) x 11(H) in.	
Platform	45(W) x 45(D) cm / 17.8(W) x 17.8(D) in.	

3. Operation

► Booting

Press the power switch into the ON position, the display illuminates and the booting screen is displayed for 2 seconds before showing the “Basic View” screen.

NOTE: If the shaker has been powered off before the completion of a run, the unit will automatically power on to the previous settings.



► Basic view screen

Ⓐ HEAT LED: indicates if the heater is activated.



Ⓑ RUN LED: indicates if the shaker is running.

Ⓒ Time setting: Indicates the set or remaining time.
(To adjust, TOUCH the time value.)

Ⓓ Temp setting: Displays the current temperature.
To set the desired temperature, TOUCH the temperature value.


Ⓔ RPM setting: Displays the set or current RPM
To set the desired shaking speed, TOUCH the RPM value.

Ⓕ Alarm ON/OFF: Alarm rings according to the condition
TOUCH to toggle the alarm on or off.

→  Alarm ON →  Alarm OFF (mute)

Ⓖ Menu : TOUCH to activate the main menu



Ⓗ Shaker Start/Stop Button
TOUCH to Start or to stop the shaker.




Ⓢ Shaker end: Indicates that time has expired and the shaker has stopped



Temperature, RPM setting change


- Ⓐ TOUCH the temperature value, The numbers become red.
 - Ⓑ "SET MODE" is displayed (flashes).
 - Ⓒ Input the desired value by pressing UP(▲)/DOWN(▼) buttons
- NOTE:** Press and hold the buttons to increase or decrease the value quickly.

- Ⓓ Apply: To store the newly set value TOUCH  (Or wait 5 seconds for the value to store automatically)
- Ⓔ Exit: If you press  after changing any values, the values are not saved and the screen returns to the previous screen.





- Ⓐ TOUCH the RPM value, The numbers become red.
 - Ⓑ "SET MODE" is displayed (flashes).
 - Ⓒ Input the desired value by pressing UP(▲)/DOWN(▼) buttons
- NOTE:** Press and hold the buttons to increase or decrease the value quickly.


- Ⓓ Apply: To store the newly set value TOUCH  (Or wait 5 seconds for the value to store automatically)
- Ⓔ Exit: If you press  after changing any values, the values are not saved and the screen returns to the previous screen.






Time setting change

- Ⓐ TOUCH the time value, The numbers become red.
 - Ⓑ "SET MODE" is displayed (flashes).
 - Ⓒ Input the desired value by pressing UP(▲)/DOWN(▼) buttons
- NOTE:** Press and hold the buttons to increase or decrease the value quickly.


- Ⓓ Apply: To store the newly set value TOUCH  (Or wait 5 seconds for the value to store automatically)
- Ⓔ Exit: If you press  after changing any values, the values are not saved and the screen returns to the previous screen.




NOTE: The time cannot be adjusted during a run. To change the time during a run, you must TOUCH . Change the time value and press  and .

The end of setting time

- ③ After the set time is over, time is displayed as 0:00:00
- ④ Shaker is stopped and the message "Shaker end" is displayed.
- ⑤ Heating LED continuously runs until the STOP button is pressed to keep the set temperature.

NOTE: The Heating is maintained even after the set time has expired to keep protecting samples at the set temperature.
If you would like to stop the heating, TOUCH .





► Advanced mode



To switch from the standard BASIC view to the ADVANCED view the shaker and heater must be off.
TOUCH "MENU" and choose the ADVANCED view.

NOTE: To return to BASIC view, TOUCH "MENU" and choose BASIC view.


BASIC ► ADVANCED








ADVANCED ► BASIC



► Temperature calibration




- ③ To enter the calibration mode enter the menu and press 
- ④ Press the "Set °C" box.
- ⑤ Input the calibrated (measured) value by using key pads and touch 
- ⑥ Apply: When you TOUCH , the changed setting values are saved and the screen turns to the previous screen.
- ⑦ If you press , the changed setting values are not saved and the screen turns to the previous screen.
- ⑧ When you press , the message is displayed as the red rectangular part



- ⑨ When you press , the calibrated value is applied and the screen turns to the MENU.
- ⑩ If you press , the calibrated value is not saved and the screen turns to MENU.
- ⑪ The last calibrated value of set °C is continuously displayed until the power is turned off and on.


※ Please measure the temperature of the inside of equipment by using a deliberate temperature measuring instrument before starting temperature calibration.
At this moment, the temperature of the inside of equipment should be stabilized. If you measure with a sensor over 1 point.

► Temperature setting (Advanced screen)





Temperature setting change

- ③ TOUCH the temperature value, the screen turns to the temperature setting screen.
- ④ TOUCH the "Temp Set" Value to adjust the set temperature.
- ⑤ Input the desired value by using the key pads in the setup screen.
- ⑥ Apply: When you press , the changed setting values are saved and the screen turns to the previous screen.
- ⑦ Exit: If you press , the changed setting values are not saved and the screen turns to the previous screen.

- ⑧ Apply: You must press  to apply new setting.


NOTE:

The temperature setting range of this instrument is 4~70°C.
When selecting a value out of this range and pressing , the message "Incorrect range" is displayed.



Adjusting the time of the temperature

- ㊤ TOUCH the temperature value, the screen turns to the temperature settings screen.
- ㊤ TOUCH the "Time Set" value to adjust the set time of the heater.
- ㊤ Input the desired value by using the key pads in the setup screen.
- ㊤ Apply: When you press **✓ Apply**, the changed setting values are saved and the screen turns to the previous screen.
- ㊤ Exit: If you press **↶ Exit** button, the changed setting values are not saved and the screen turns to the previous screen



Adjusting the alarm setting


The alarm value is the difference between the set temperature and the measured (displayed) temperature. If the difference exceeds this value (once the temperature has been reached) the alarm will trigger.

- ㊤ TOUCH the temperature value, the screen turns to the temperature settings screen.
- ㊤ TOUCH the "Alarm Set" value to enter the alarm setup.
- ㊤ Input the desired value by using the key pads in the setup screen.
- ㊤ Apply: When you press **✓ Apply**, the changed setting values are saved and the screen turns to the previous screen.
- ㊤ Exit: If you press **↶ Exit** button, the changed setting values are not saved and the screen turns to the previous screen.

※ The alarm setting range is 0.1~10.0.

When putting the value out of the range and pressing **✓ Apply**, the message "Incorrect range" is displayed.

NOTE: With a set temperature of over 50C, an alarm value of +2.0 is recommended.



Alarm off

The alarm can be deactivate in both the BASIC View and the ADVANCED View by TOUCHING the  Alarm key to toggle on or off.

► Shaker setting (Advanced)





RPM setting change

Ⓐ TOUCH the RPM Shaker value, the screen turns to the Shaker setting screen.


Ⓑ TOUCH the "RPM Set" value to adjust the set speed.


Ⓒ Input the desired value by using the key pads in the setup screen.

Ⓓ Apply: When you press , the changed setting values are saved and the screen turns to the previous screen.

Ⓔ Exit: If you press  button, the changed setting values are not saved and the screen turns to the previous screen.

※ The RPM setting range is 30~300RPM.

When putting the value out of the range and pressing , the message "Incorrect range" is displayed.





Adjusting the shaking time

Ⓐ TOUCH the RPM Shaker value, the screen turns to the Shaker setting screen.

Ⓑ TOUCH the "Time Set" value to adjust the set shaking time.

Ⓒ Input the desired value by using the key pads in the setup screen.

Ⓓ Apply: When you press , the changed setting values are saved and the screen turns to the previous screen.

Ⓔ Exit: If you press  button, the changed setting values are not saved and the screen turns to the previous screen.



Alarm setting change


The alarm value is the difference between the set speed and the measured (displayed) speed.


If the difference exceeds this value (once the RPM speed has been reached) the alarm will trigger.

Ⓐ TOUCH the RPM Shaker value, the screen turns to the Shaker setting screen.


Ⓑ TOUCH the "Alarm Set" value to adjust the set speed.

Ⓒ Input the desired value by using the key pads in the setup screen

Ⓓ Apply: When you press , the changed setting values are saved and the screen turns to the previous screen.

Ⓔ Exit: If you press  button, the changed setting values are not saved and the screen turns to the previous screen.

※ The RPM setting range is 1~30RPM.


When selecting the value out of this range and pressing , the message "Incorrect range" is displayed.



Alarm off

The alarm can be deactivate in both the BASIC View and the ADVANCED View by TOUCHING the  key to toggle on or off.

► Temperature step setting



Temperature time setting

- ③ TOUCH the temperature value to enter the temperature setup screen.
- ④ Select , the screen moves to the temperature step & setup screen.
- ⑤ You can select/link up to 4 desired temperatures. Choose each 4 desired temperatures by pressing on the value.
(Leave the time and temperature values as "0" to skip the unlinked step.)
- ⑥ Once a temperature value is selected, use the keypad in the setup screen to select the desired temperature.
- ⑦ Apply: When you press , the changed setting values are saved and the screen turns to the previous screen.
- ⑧ Exit: If you press , the changed setting values are not saved and the screen turns to the previous screen.
- ⑨ Once the desired temperatures have been chosen, adjust the time values by following the same steps.
- ⑩ There are 3 modes available for Step Operation : "One cycle", "Repeat Cycle", "Step off"

NOTE:

Repeat cycle will result in continue repeating of all programmed steps until the stop button is pressed.

► Shaker step setting



RPM time setting


- ③ TOUCH the RPM shaker value to enter the shaker setup screen.
- ④ Select **Step**, the screen moves to the shaker step & setup screen.
- ⑤ You can select/link up to 4 desired shaking speeds. Choose each 4 desired speeds by pressing on the speed value.
(Leave the time and speed values as "0" to skip the unlinked step.)
- ⑥ Once a RPM step is selected, use the keypad in the setup screen to select the desired RPM value.
- ⑦ Apply: When you press **Apply**, the changed setting values are saved and the screen turns to the previous screen.
- ⑧ Exit: If you press **Exit** button, the changed setting values are not saved and the screen turns to the previous screen.
- ⑨ Once the desired RPM speeds have been chosen, adjust the time values by following the same steps.
- ⑩ There are 3 modes available for Step Operation : "One cycle", "Repeat Cycle", "Step off"

NOTE:

Repeat cycle will result in continue repeating of all programmed steps until the stop button is pressed.

► Step cycle ending screen

- ④ When the time for each step has expired, the “End” message is displayed.
- ⑤ “Set” displays the setting value of 1st step (Default value is displayed when Step Off)
Time is displayed as 0.
- ⑥  or  is displayed as the instrument can be restarted at any time.



4. Additional Operating Instructions

► H2012 (Cooling) Model Only:

If the instrument will be set to sub ambient temperatures, please turn on the “Cooling Switch” located on the front left side of the instrument.

DO NOT power on the “Cooling Switch”, if the set temperature is more than 5°C above the room temperature.

※ Powering on this switch when the temperature is set +5°C above ambient may cause damage to the cooling system.



▲ Compressor Switch (i.e. “Cool Switch”)

► Over Temperature Protection:

The Incu-Shaker 10 Series includes an Over Temperature Safety Knob on the right side of the instrument. This safety feature results in an automatic shut-off once the measured temperature inside of the chamber has exceeded the set point of the knob.

When the automatic shut-off has been activated, the SAFETY LED on the front panel will be illuminated. The default setting for this safety knob is 70°C. The value of this knob can easily be adjusted by the user.

NOTE: This safety knob should always be set at least 5°C higher than the programmed temperature on the digital display. If not, the incubator may be prevented from reaching the set temperature.



▲ Over Temperature Safety Knob

► Platform Installation:

Microplate Platform and Spring Platform Installation:

The Incu-Shaker 10L(10LR) includes a flat non-slip platform for low speed operation with many common vessels, including trays and dishes. For installation of any of the optional platforms the user must:


- 1) Unscrew the four plastic screws from the corners of the internal platform and remove the green rubber mat.
- 2) Clean any dirt or debris from the metal base-platform and apply the desired optional platform so that the four corner holes are aligned.
- 3) Screw in the four platform screws tightly by hand.

Magic Clamp Platform Installation:

If you have ordered the H1010-MR MAGic clamp platform as an accessory with your Incu-Shaker 10L(10LR), the platform may arrive installed on the shaker. If not, please follow the steps below for installation instructions.

- 1) Unscrew the four corner thumb screws from the internal platform and remove the green rubber mat.
- 2) Clean any dirt or debris from the metal base platform and apply the MAGic Clamp platform so that the four corner holes are aligned.
- 3) Screw in the four platform screws tightly by hand.

5. Warning Messages & Fuse Replacement




- The fuse is stored in the fuse box of the integral power inlet on the backside of the instrument.
- In the case you need to use the spare fuse, put a 7-shaped tool, a flat-head screwdriver or a tweezers in the little U-shaped hole on the bottom of fuse box to open the cover, remove the installed fuse and replace it with the included spare.

► Event view

Door open message

When the door has been left open for opening the door, the message “Door” is displayed on the basic and advanced screen.

NOTE: Repeat cycle will result in continue repeating of all programmed steps until the stop button is pressed.




Temperature/Shaker Alarm

The message “low’ or “high” is displayed on the display when the measured value is deviating from the set value.

The message is only displayed when the difference exceeds the user programed value for the alarm.

NOTE: The message is cleared when the equipment maintains the values for 1 minute following the activation of the alarm.



Maintenance:

The platform of the Incu-Shaker 10L (10LR) should be cleaned regularly with a soft cloth dampened with water or alcohol. (Always ensure that the machine has been powered down prior to cleaning.)

In the event of a spillage:

Power down the unit, remove the platform and dry the internal chamber with a soft cloth.

DO NOT attempt to operate the unit until the spillage had been properly dried.

Service and Contact

Service on the Incu-Shaker 10L (10LR) should only be performed by qualified service personnel. To request service or technical support, please contact Benchmark Scientific or your local benchmark Scientific representative.

Ph: +1-908-769-5555
Fax: +1-732-313-7007
Em: Info@BenchmarkScientific.com
Web: www.BenchmarkScientific.com
P.O. Box 709 Edison, NJ 08818, USA

